

**SCREENING SYSTEM FOR HEART VALVE  
DISEASE**

**MOHD ZUBIR BIN SUBOH**

**UNIVERSITI MALAYSIA PERLIS  
2011**

© This item is protected by original copyright



## **Screening System for Heart Valve Disease**

By

**Mohd Zubir Bin Suboh  
(0731310194)**

A thesis submitted in fulfillment of the requirements for the degree of Master  
of Science (Mechatronic Engineering)

**School of Mechatronic Engineering  
UNIVERSITI MALAYSIA PERLIS**

**2011**

## ACKNOWLEDGEMENTS

Alhamdulillah, thanks to Allah S.W.T for his blessing and guidance for me to finish this project and thesis successfully. The thesis has been realized by the combined work and skills of many people. Firstly I would like to thank the Ministry of Science, Technology and Innovation Malaysia (MOSTI) for funding the research and giving scholarship during the research periods.

My special thanks goes to Prof. Dr. Mohd Yusoff bin Mashor for his time and efforts to advise, guide, support and encourage me to finish the work. Then, I would like to express my sincere gratitude to my co-supervisors, Assoc. Prof. Abd. Rahman Mohd Saad (School of Computer Engineering) and Dr. Mohd Sapawi Mohamed (Department of Cardiology, Hospital Universiti Sains Malaysia (HUSM)) for their guidance, support, comments and advises. I would also like to thank Dr. Khor Boon Tat, cardiologist from Hospital Tuanku Fauziah (HTF) and his staff for their cooperation in providing guidance and support during data collection.

I also want to thank all the members of Autonomous and Machine Vision (AutoMAV) Cluster who have helped me one way or another to remove the obstacles and difficulties faced in this research. Special thanks are given to Mr. Mohd Shaiful Aziz bin Rashid Ali and Mr. Muhyi bin Yaakob for their help in developing the embedded system.

Big thanks go to my parents and family for their endless advice, love and prayers. My thanks and appreciation goes to all these people and others who were directly or indirectly involved in this research. Without them, this research work would never have been completed.

## TABLE OF CONTENTS

	<b>PAGE</b>
<b>ACKNOWLEDGEMENTS</b>	ii
<b>TABLE OF CONTENTS</b>	iii
<b>LIST OF FIGURES</b>	viii
<b>LIST OF TABLES</b>	xi
<b>LIST OF ABBREVIATIONS</b>	xiii
<b>ABSTRAK</b>	xvi
<b>ABSTRACT</b>	xviii
<b>CHAPTER 1 INTRODUCTION</b>	
1.1 Introduction	1
1.2 Problem Statements	2
1.3 Objective of Research	5
1.4 Scope of Research	6
1.5 Thesis Outline	7
<b>CHAPTER 2 LITERATURE REVIEW ON HEART VALVE DISEASE AND HEART SOUND ANALYSIS</b>	
2.1 Introduction	8
2.2 Anatomy of the Heart	8
2.3 Heart Valve Disease	10
2.3.1 Type of Heart Valve Disease	11
2.3.2 Symptoms	11

2.3.3	Diagnosis Techniques	12
2.3.3.1	Heart Sound Auscultation	13
2.3.3.2	Phonocardiography (PCG)	13
2.3.3.3	Electrocardiography	14
2.3.3.4	Chest X-Ray	14
2.3.3.5	Cardiac Characterization	15
2.3.3.6	Stress Test	15
2.3.3.7	Echocardiography	15
2.3.3.8	Magnetic Resonance Imaging (MRI)	16
2.4	Heart Sound Auscultation	17
2.5	Heart Sound	19
2.5.1	First Heart Sound	20
2.5.2	Second Heart Sound	21
2.5.3	Third Heart Sound	21
2.5.4	Fourth Heart Sound	22
2.5.5	Heart Murmur	22
2.6	Heart sound Analysis	23
2.6.1	Segmentation of Heart Sound	24
2.6.2	Feature Extraction	25
2.6.3	Classification of Heart Sound	29
2.7	Existing System	32
2.8	Summary	34

## **CHAPTER 3     METHODOLOGY**

3.1 Introduction	35
3.2 Data Collection Procedures of the Heart Sound Signal	35
3.2.1 Data Collected from Heart Sound Manipulator Software	36
3.2.2 Recorded Data from Hospital Tuanku Fauziah (HTF)	38
3.2.3 Data Collected from Internet	39
3.3 PC Based System Development	41
3.3.1 Pre-Processing of Heart Sound Signal	42
3.3.2 Segmentation of Heart Sound Signal	43
3.3.3 Feature Extraction of Heart Sound Signal	59
3.3.4 Classification of Heart Sound Signal	62
3.3.5 Complete PC Based Screening System	69
3.4 Implementation of the Screening System in Embedded Board	70
3.5 Embedded System Development	70
3.5.1 Electronic Stethoscope	71
3.5.2 Single Board Computer (SBC)	74
3.5.3 VC21PC1 Portable Computing COG	76
3.5.3.1 Color TFT (Thin Film Transistor) LCD	78
3.5.3.2 Encoder and Decoder (CODEC)	78
3.5.3.3 VC21PC1 Input Switches	80
3.5.4 Embedded Board Specification	80
3.5.5 Firmware Development	81
3.5.6 Graphical User Interface (GUI) Design	83

3.5.6.1	Record and Play Function	84
3.5.6.2	Phonocardiogram (Plot of Heart Sound Signal)	84
3.5.6.3	Result and Analysis	85
3.5.6.4	Details Results for Each Sample	85
3.5.6.5	Reset and Clear Screen	86
3.5.7	Simplification of Functions and Algorithms	87
3.6	Summary	92
<b>CHAPTER 4     RESULTS AND DISCUSSION</b>		
4.1	Introduction	93
4.2	Heart Sound Analysis	94
4.2.1	Heart Sound Segmentation	94
4.2.2	Cross-Correlation Analysis	101
4.2.2.1	Samples with Different Diseases	102
4.2.2.2	Samples with Different Durations	103
4.2.2.3	Samples with Different Cycle Sequences	104
4.2.2.4	Samples with Different Amplitudes	107
4.2.3	Heart Sound Classification	108
4.2.3.1	Classification of Heart Sound Signals using Standard MLP Network	110
4.2.3.2	Classification of Heart Sound Signals using Hierarchical MLP Network	115
4.3	Accuracy of the Complete System	120
4.3.1	Classification Accuracy on PC Platform	122

4.3.2 Classification Accuracy on Embedded Platform	126
4.4 Comparison with Other Existing Systems	130
4.5 Conclusion	133
<b>CHAPTER 5 CONCLUSIONS</b>	
6.1 Conclusion	135
6.2 Contribution of Research	137
6.3 Problems Encountered during Research Period	138
6.4 Suggestion for Future Works	139
<b>REFERENCES</b>	140
<b>APPENDIX A</b> Heart Sound Data Sets	152
<b>APPENDIX B</b> Schematics of the Embedded System	159
<b>APPENDIX C</b> Approximation Technique using Lookup Tables	169
<b>APPENDIX D</b> Feature Extraction and Classification Results	172
<b>LIST OF PUBLICATIONS</b>	187

## LIST OF FIGURES

		<b>PAGE</b>
Figure 2.1	Human Heart Illustration ( <i>Health Sciences Centre, 2010</i> )	9
Figure 2.2	A Normal Adult 12-Leads ECG Signals ( <i>Review &amp; Study Guide Online (PTORG), 2008</i> )	14
Figure 2.3	Frequency Response of a Mechanical Stethoscope ( <i>Ertel, Lawrence, Brown &amp; Stern, 1966</i> )	18
Figure 2.4	Actual Heart Valve Positions Together With Auscultation Positions ( <i>Human, 2006</i> )	19
Figure 2.5	A cardiac cycle with S1, S2, S3 and S4	20
Figure 3.1	Heart Sound Manipulator	37
Figure 3.2	Heart's Shape and Heart Sound Signals of a Patient at Three Different Positions ( <i>Stein &amp; Eisenberg, 1998</i> )	39
Figure 3.3	Block Diagram of PC Based Screening System	41
Figure 3.4	Heart Sound Envelopgrams	46
Figure 3.5	Effects on Peak Selections on Different Threshold Ranges	48
Figure 3.6	(a) One Peak is Selected in One Cycle (b) Two Peaks are Selected in One Cycle	49
Figure 3.7	Example of Peak Interval and Interval Consistency Level on One Peak selected in One Cycle	52
Figure 3.8	Example of Systolic and Diastolic Intervals and with their Consistency Level on Two Peaks Selected in One Cycle	53
Figure 3.9	(a) Missing Peak (b) Unwanted Peak	54
Figure 3.10	Flow Chart of Automated Heart Sound Segmentation	58
Figure 3.11	Power Spectrum of Reference Heart Sound Sample	61
Figure 3.12	Cross-Correlation Plot for Five category of Heart Sound	62

Figure 3.13	Selecting 500 Significant Point from the Cross-correlation Plot	63
Figure 3.14	Similar Final Plot Pattern of the Cross-correlation Plot	64
Figure 3.15	Selecting 700 Significant Point from the Cross-correlation Plot	65
Figure 3.16	Classification of 5 Categories of Heart Sound Signal using Standard MLP Network	66
Figure 3.17	Classification of 5 Categories of Heart Sound Signal using Hierarchical MLP Network	68
Figure 3.18	Log-Sigmoid Transfer Function	68
Figure 3.19	Screening System Interface Developed in PC Platform	69
Figure 3.20	Block Diagram for Embedded Screening System	71
Figure 3.21	Welch Allyn Master Elite Plus Electronic Stethoscope	73
Figure 3.22	VCMX212 SBC	76
Figure 3.23	Top and Bottom View of VC21 Multimedia Computing Cogs	77
Figure 3.24	Compilation Stages for GNU C Compiler	82
Figure 3.25	GUI Design for the Embedded System	83
Figure 3.26	Plot of Heart Sound Signal in Phonocardiogram Window	84
Figure 3.27	Details Results of Each Segmented Sample	86
Figure 3.28	Complex Number of Modulus One and Phase $\theta$ Represented as a Vector on a Unit Circle	87
Figure 3.29	90 Values of a Lookup Table for Calculating Sine and Cosine Function Values	88
Figure 3.30	Four Different Presentation of the Lookup Table to Generate Output of Sine and Cosine Function for 1 to 360 of Angles	89
Figure 3.31	Sigmoid Function	90

Figure 4.1	Good Segmentation Result	97
Figure 4.2	Poor Segmentation Result	98
Figure 4.3	Cross-Correlation Results of Sample with Different Disease	102
Figure 4.4	Cross-Correlation Results of Samples with Different Duration	104
Figure 4.5	Cross-Correlation Results of Normal Samples with Different Cycle Sequence	105
Figure 4.6	Cross-Correlation Results of Abnormal Samples with Different Cycle Sequence	106
Figure 4.7	Cross-Correlation Results of Samples with Different Amplitudes of Murmurs	107

© This item is protected by original copyright

## LIST OF TABLES

		<b>PAGE</b>
Table 2.1	Valve Locations and Auscultation Positions	19
Table 3.1	List of Heart Sound Signal Recorded from Heart Sound Manipulator	37
Table 3.2	List of Heart Sound Signal Recorded from HTF	40
Table 3.3	List of Heart Sound Signal collected from the Internet	41
Table 3.4	Effect on Peaks Selection on Different Threshold Ranges	48
Table 3.5	Automated Peak Selection Based on the Number of Peak in 3 seconds Heart Sound Signal	51
Table 3.6	Absolute Maximum Rating	83
Table 3.7	Recommended Operating Range	83
Table 3.8	Mechanical Specification	83
Table 3.9	Procedure to calculate Sine and Cosine for 360 degrees using SinKos Lookup Table	91
Table 3.10	PLAN Approximation Algorithms	93
Table 4.1	Segmentation's Result in PC Platform	95
Table 4.2	Segmentation's Result in Embedded Platform	100
Table 4.3	List of Samples with Its Duration	103
Table 4.4	List of Normal Samples with Its Duration and Starting Point of the Cycle	105
Table 4.5	List of Abnormal Samples with Its Duration	106
Table 4.6	Group of Data Set for Heart Sound Signals Classification	109
Table 4.7	Number of Samples for the Standard MLP Training and Testing	111

Table 4.8	Classification Accuracy of 5 Categories of Heart Sound Signals after 10 Trials at 30 Different Hidden Neurons using the First Method on First Data Set	113
Table 4.9	Classification Accuracy of 5 Categories of Heart Sound Signals after 10 Trials at 30 Different Hidden Neurons using the First Method on Second Data Set	114
Table 4.10	Number of Samples for the First Network Training and Testing	115
Table 4.11	Classification Accuracy of 2 Categories of Heart Sound Signals after 10 Trials at 30 Different Hidden Neurons using the First Network of the Second Method	116
Table 4.12	Number of Samples for the Second Network Training and Testing	118
Table 4.13	Classification Accuracy of 4 Categories of heart sound signals after 10 Trials at 30 Different Hidden Neurons using the Second Network of the Second Method	119
Table 4.14	Final Output of Complete System Classification	120
Table 4.15	Screening Result of the PC Based System	124
Table 4.16	Classification Result of the PC Based System	125
Table 4.17	Screening Result of the Embedded System	128
Table 4.18	Classification Result of the Embedded System	129
Table 4.19	Comparison of Heart Valve Disease Screening System with Other Existing System	131

## LIST OF ABBREVIATIONS

ADC	Analog to Digital Converter
AGC	Automatic Gain Control
ANN	Artificial Neural Network
AR	Aortic Regurgitation
ARM	Auto Regression Model
AS	Aortic Stenosis
ASD	Atrial Septal Defect
AWGN	Additive White Gaussian Noise
BPM	Beats per Minute
CAD	Coronary Artery Diseases
CODEC	Coder and Decoder
CPM	Cycle per Minute
CPS	Cycle per Second
CT	Computed Tomography
CVDs	Cardiovascular Diseases
CWT	Continuous Wavelet Transform
DA	Discriminant Analysis
DFT	Discrete Fourier Transform
DMA	Direct-Memory-Access
DSP	Digital Signal Processor
DWT	Discrete Wavelet Transform
ECHO	Echocardiograph
EGM	Egeneral Medical
EKG or ECG	Electrocardiogram
ESD	Electro-Static Discharge
FFT	Fast Fourier Transform
FIFO	First In First Out
FPGA	Field-Programmable Gate Array
GPIO	General Purpose Input/Output

GUI	Graphical User Interface
HTF	Hospital Tuanku Fauziah
HVD	Heart Valve Disease
I/O	Input and Output
JRE	Java Runtime Environment
KMD	Kuwait Medical Department
kNN	K-Nearest Neighbour
ksps	Kilo Sample Per Second
LCD	Liquid Crystal Display
Li-ION	Lithium-ION
MFCC	Mel Frequency Cepstrum Coefficient
MLP	Multi-Layer Perceptron
MMC	Multi-Media Card
MoH	Ministry of Health
MR	Mitral Regurgitation
MRI	Magnetic Resonance Imaging
MS	Mitral Stenosis
N	Normal Heart Sounds
OS	Operating System
PC	Personal Computer
PCB	Printed Circuit Board
PCG	Phonocardiograph
PDA	Personal Digital Assistant
PGA	Programmable Gain Amplifier
PR	Pulmonary Regurgitation
PS	Pulmonary Stenosis
PWM	Pulse Width Modulator
RISC	Reduced Instruction Set Computing
SBC	Single Board Computer
SFZ	Short-Time Fourier Analysis
SNR	Signal to Noise Ratio

SPI	Serial Peripheral Interfaces
STFT	Short Time Fourier Transform
S1	First Heart Sound
S2	Second Heart Sound
S3	Third Heart Sound
S4	Fourth Heart Sound
TCE	The Cardiac Exam
TEE	Transesophageal Echocardiogram
THI	Texas Heart Institute
TR	Tricuspid Regurgitation
TRAINLM	Levenberg-Marquardt training
TS	Tricuspid Stenosis
TSC	Thinklab Stethoscope Community
UART	Universal Asynchronous Receiver/Transmitters
UOW	University of Washington
VSD	Ventricular Septal Defect
WHO	World Health Organization
WT	Wavelet Transform

## Sistem Saringan Untuk Penyakit Injap Jantung

### ABSTRAK

Teknik pertama yang digunakan oleh doktor dan kardiologi untuk mengesan penyakit jantung adalah teknik auskultasi terhadap bunyi jantung. Namun begitu, satu jangka masa yang panjang diperlukan bagi memperolehi kemahiran auskultasi tersebut. Menyedari pentingnya teknik auskultasi, kajian ini dilakukan untuk membina satu sistem saringan yang boleh mengenalpasti bunyi jantung normal dan abnormal serta mengelaskan bunyi jantung abnormal tersebut kepada empat kategori umum penyakit injap jantung. Penyakit-penyakit tersebut adalah *Aortic Regurgitation* (AR), *Aortic Stenosis* (AS), *Mitral Regurgitation* (MR) dan *Mitral Stenosis* (MS). Sistem saringan ini boleh memproses pensegmenan, pengekstrakan ciri dan pengelasan isyarat bunyi jantung secara automatik. Proses pensegmenan yang diperkenalkan dalam sistem ini adalah berasaskan informasi masa dari isyarat bunyi jantung. Ia digunakan dalam penghasilan sampel untuk pengekstrakan ciri-ciri isyarat bunyi jantung. Sampel dari domain masa ini ditukarkan ke dalam domain frekuensi dan spektrum kuasa bagi sampel ini dikira. Spektrum kuasa ini pula akan digunakan untuk mendapatkan ciri-ciri isyarat bunyi jantung dengan menggunakan kaedah kolerasi silang. Kaedah kolerasi silang yang dicadangkan ini adalah sangat teguh di mana tempoh masa sampel, jujukan kitaran sampel dan kekuatan isyarat bunyi jantung tidak memberikan kesan ketara kepada spektrum kuasa itu sendiri. Sampel yang telah diekstrakkan ini kemudiannya akan dikelaskan dengan menggunakan sistem Rangkaian Perseptron Berbilang Lapis biasa dan Rangkaian Perseptron Berbilang Lapis berhierarki. Keputusan kajian menunjukkan pengelasan isyarat bunyi jantung yang menggunakan Rangkaian Perseptron Berbilang Lapis berhierarki memberikan ketepatan pengelasan sebanyak 100%, lebih baik daripada Rangkaian Perseptron Berbilang Lapis biasa iaitu 85.71%. Ini disebabkan kekompleksan dalam pengelasan 5 isyarat bunyi jantung berbeza dapat dikurangkan dengan membahagikan 5 kelas bunyi jantung itu kepada dua bahagian dengan menggunakan Rangkaian Perseptron Berbilang Lapis berhierarki. Sistem lengkap yang terdiri daripada proses pensegmenan, pengekstrakan ciri serta pengelasan isyarat bunyi jantung dibina di dalam sistem berasaskan komputer dan seterusnya diimplementasikan ke dalam satu sistem terbenam. Sistem terbenam ini terdiri daripada elektronik stetoskop, papan multimedia lengkap (VC21PC1) dan papan komputer tunggal (VCMX212) sebagai terasnya. Keberkesanan sistem berasaskan komputer dan sistem terbenam ini akan dibincangkan di dalam kajian ini. Sebanyak 646 sampel dari 39 subjek telah digunakan di dalam kajian ini. Hasil kajian menunjukkan bahawa kedua-dua sistem ini, masing-masing menghasilkan 96.3%, 92.59% dan 94.44% untuk kekhususan, kepekaan dan ketepatan saringan bunyi jantung normal dan abnormal. Ini menunjukkan kaedah yang dicadangkan adalah baik dan boleh dipercayai. Walaubagaimanapun, untuk klasifikasi yang lebih khusus, sistem berasaskan komputer telah menghasilkan ketepatan sebanyak 94.44% manakala sistem terbenam dengan ketepatan 87.04% sahaja. Hal ini berlaku disebabkan beberapa penghampiran yang digunakan di dalam pengiraan spektrum kuasa dan keluaran bagi pengelas Rangkaian Perseptron Berbilang Lapis. Perbandingan juga turut dibuat dengan sistem-sistem yang sedia ada dan didapati bahawa sistem yang dicadangkan ini

telah menghasilkan ketepatan saringan yang standing (94.44%) untuk pengelasan bunyi jantung normal dan abnormal. Ketepatan untuk pengelasan khusus bagi penyakit injap jantung adalah secara amnya lebih baik (87.04%) berbanding sistem-sistem sedia ada yang lain. Perbandingan secara langsung tidak dapat dilakukan kerana data dan kaedah yang digunakan oleh penyelidik-penyelidik yang lain adalah sama sekali berbeza.

© This item is protected by original copyright

## Screening System for Heart Valve Disease

### ABSTRACT

The first method applied by physician and cardiologists to detect heart disease is heart sound auscultation. However the skill of auscultation takes many years to acquire. Acknowledging the importance of heart sound auscultation, this research is conducted to develop a screening system that can classify normal and abnormal heart sound as well as categorizing the abnormal heart sound signal into four common categories of heart valve disease. The diseases are Aortic Regurgitation (AR), Aortic Stenosis (AS), Mitral Regurgitation (MR) and Mitral Stenosis (MS). The screening system is able to perform automated process of segmentation, feature extraction and classification of the heart sound signals. The segmentation process that is based on the time properties of the heart sound is introduced in this study to produce samples for feature extraction. The sample is converted to frequency domain and power spectrum of the signal is calculated. Power spectrum of the signal is used to get the heart sound features using cross-correlation method. The proposed method is a robust method where samples duration, cycle sequence and amplitude of the heart sound and murmur are not significantly affecting the power spectrum itself. The extracted frequency features are classified using standard Multi-Layer Perceptron (MLP) network and hierarchical Multi-Layer Perceptron network. Classification accuracy obtained from hierarchical MLP network is 100%, better than standard MLP network with accuracy of 85.71%. This is due to the complexity in classification of 5 types of heart sound signals has been reduced to two parts by using Hierarchical MLP network. A complete system that includes the process of segmentation, feature extraction and segmentation of the heart sound signal is developed in PC based platform and implemented in an embedded system. The embedded system is consists of electronic stethoscope, multimedia board (VC21PC1) and a single board computer (VCMX212) as the core. Efficiency of both PC based and embedded system is investigated in this study. A total of 646 samples from 39 subjects have been used in this study. The results show that both PC based and embedded system has produced 96.3%, 92.59% and 94.44% of screening specificity, sensitivity and accuracy for normal and abnormal classification, respectively. This showed that the proposed method is good and reliable. However, for specific classification on the other 4 type of abnormal heart sound signal, the PC based system has produced 94.44% accuracy while the embedded system only produced 87.04% accuracy. The reason is that a few approximations were applied in calculating the features and output of the MLP network. Comparison is also made with other existing systems and it is found that the proposed system has produced a comparable screening accuracy (94.44%) for normal and abnormal classification and generally better accuracy for heart valve diseases classification (87.04%). Direct comparison cannot be made because the data and method used by the other researchers are totally different.

# CHAPTER 1

## INTRODUCTION

### 1.1 Introduction

Heart disease is a general term for a variety of different diseases affecting the heart. Heart disease is a type of cardiovascular diseases (CVDs) which is the number one cause of death globally. A report from World Health Organization (WHO) in 2009 stated that an estimated of 17.5 million people died because of CVDs in 2005. More people die annually from CVDs than any other cause. The organization expected that by 2030, almost 23.6 million people will die from CVDs, mainly from heart disease and stroke. These are projected to remain the single leading cause of death. The largest percentage increase will occur in the Eastern Mediterranean Region. The increase in number of deaths will occur in South-East Asian Region (*World Health Organization [WHO], 2009*).

National Vital Statistic (2009) reported that heart disease was the first cause of death globally with 26 % of total death. The same scenario happens in Malaysia. Statistics provided by Malaysia Ministry of Health (MOH) claimed that the first principal cause of death in MoH hospital is heart diseases and diseases of pulmonary circulation with 16.09% (*Ministry of Health Malaysia, 2010*). Therefore the study of normal and pathological heart behavior became an active research area. In particular, the study of the shape and motion of the heart is important because they are very much affecting the heart sound signal. Many heart diseases are strongly correlated to these two factors (Park, Montillo, Metaxas & Leon, 2005).

Heart auscultation and diagnosis are quite complicated, depending not only on the heart sound but also on other factors such as the acquisition method and patient condition (Stasis, Loukis, Pavlopoulos & Koutsouris, 2004). In the last two decades, many research activities were conducted concerning automated and semi-automated heart sound diagnosis. The researches were concentrated at three major tasks which are segmentation of the heart sound, feature extraction and classification of heart sound using artificial intelligence system. In this study, a screening system that includes the three major tasks is proposed to screen heart sound signal by using an electronic stethoscope. The screening system was developed on two platforms, personal computer (PC) and embedded system platform. The system consists of segmentation, feature extraction and classification of the heart sound signals. An artificial intelligence system is used in this study to classify normal and abnormal heart sound signals as well as classify the abnormal heart sound signals into four common categories of heart sound from Heart Valve Disease (HVD).

## **1.2 Problem Statements**

The problems that motivate this study are limitation of heart sound auscultation technique, lack of intelligence utilization in phonocardiograph instrument, unavailability of advance diagnostic equipments and necessity of robust and reliable technique in signal processing.

### **A. Limitation of Heart Sound Auscultation Technique**

Heart sound analysis has been an essential clinical skill for every doctor and medical student. As a subjective and qualitative method, heart sound analysis by auscultation depends highly on the skills and experience of the listener. It takes many years of experience before one can master the skill of auscultation. Different auscultator might describe the heart sound differently because the heart sound especially low murmurs are very hard to be heard even with the assists of electronic stethoscope (Voss, Herold, Schroeder, Nasticzky, Mixli, Ullrich, & Huebner, 2003). Despite the importance of auscultation, trainees are not adequately trained in their training program to master the skill of auscultation (Criley, 2000). This is very important matter to dealt with because poor performance given by physicians and cardiologists to recognize the significant cardiac lesions by physical examination might leads to adverse outcomes to the patient as well as unnecessary costs for inappropriate and even potentially hazardous laboratory test. This is why the research is very important in order to replace the auscultation technique using analytical perspective. It will be very helpful to doctors and physician to screen the heart sound signal effectively.

### **B. Lack of Intelligence Utilization in Phonocardiograph Instrument**

The phonocardiogram is a plot of high fidelity recording of the heart sounds and murmurs using a machine called phonocardiograph. This plot is very important since it is proved to be very useful in understanding the heart sound signal and has high potential to detect heart disease (Rangayyan & Lehner, 1988). However, the old analogue phonocardiograph is not having quantitative analysis capability. Digital phonocardiograph adopted traditional stationary analysis and time-frequency analysis such as Short-time

Fourier Analysis (SFZ), Auto Regression Model (ARM) and Wavelet Transform. These methods were used to analyse first and second heart sound and resulted in some achievements (Zehan, Shiyong, Li, Yuli, & Shouzhong, 1998). Nevertheless, in many of these systems, since artificial intelligence technology was not introduced, the difficulties of identifying some components of abnormal heart sounds cannot be properly overcome and the number of diseases which are classified is quite less (Wenhui, Duanrong, Yun & Zhifen, 1997). Hence, it is very advantageous if one can design an instrument that not only can show the phonocardiogram signal but also with the ability to classify the heart sound signals using an artificial intelligence system.

### **C. Unavailability of Advance Diagnostic Equipments in Primary Healthcare**

Advanced imaging techniques such as Electrocardiogram (EKG or ECG), Echocardiograph (ECHO), Magnetic Resonance Imaging (MRI) and Computed Tomography (CT) can provide more direct accurate evidence of heart disease better than heart auscultation. Conversely, these modalities are normally available at advanced hospitals because they are costly, large in size and operationally complex. Consequently it is not suitable for use in rural areas, in homecare and generally in primary healthcare setups (Javed, Venkatachalam, and Ahmad Fadzil, 2006). Therefore in numerous cases the heart sound diagnosis is the possible economical and quick alternative to detect the heart condition under emergency conditions. For example, patients will not have to be transferred to diagnostic room or to another room that is near to the diagnostic room, just to examine their heart condition frequently if a portable device can be brought to them and do the same task.

#### **D. Necessity of Robust and Reliable Technique in Signal Processing**

The application of conventional and advanced digital signal processing techniques to the analysis of heart sounds requires computer-based capabilities, necessitating the use of relatively powerful processors, which are found in desktop or laptop computers. This is because the processing involves complex mathematical algorithm that can only easily be solved using dedicated function, toolbox or library provided by other simulator software such as MATLAB and LabVIEW. The system might requires high processing speed of processor and high memory size for data storage and calculation. Hence the need of a robust and reliable analysis of heart sounds is not only justified but highly desirable (Brusco & Nazeran, 2004).

#### **1.3 Objective of Research**

The objectives of this study are:

1. To develop a technique to extract some features of heart sound for heart valve disease.
2. To build an automated system to do heart sound segmentation, feature extraction and classification of heart sound signals in PC and embedded platforms.
3. To investigate the efficiency of the proposed screening system in PC and embedded platforms.